



Hinge Axis Location Concept:

- On the patient's RIGHT side, opening movement creates a clockwise rotation of the mandible. Therefore, the true hinge axis will always be located 90 degree from the direction the stylus moves in opening movement.
- On the patient's LEFT side, opening movement creates a counter-clockwise rotation of the mandible. Therefore, the true hinge axis will always be located 90 degree from the direction the stylus moves in opening movement.